Hydroclave Maintenance Schedule

2030 E. University Drive
Rancho Dominguez, CA

Phone: 310-328-6600 Fax: 310-603-9625

✔ *recommended service*

Recommended by Thermal Equipment Corporation

ASME No.	*Month*	January				February				March				April				May				June				July				August				September				October				November				December			
Year:	*Week*	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4	1	2	3	4
Hydroclave Closure																																																	
Inspect sliding surfaces of door	1 month	✔				✔				✔				✔				✔				✔				✔				✔				✔				✔				✔				✔			
Strip sliding surfaces of lubricant and re-lubricate	1 year	✔																																														
Inspect split ring angle	1 year	✔																																														
Inspect interlocks operations and functions	1 month	✔				✔				✔				✔				✔				✔				✔				✔				✔				✔				✔				✔			